



California Medical Device Recall Information



Recall Name

Honeywell Recalls Eyesaline Eyewash Solution Due to Potential Microbial Contamination

Recall Date	Product Description	Recalling Firm	Recall Reason
8/19/16	Eyesaline Eyewash Solution, 32 oz.	Honeywell, Inc. Morris Plains, NJ	<i>Potential for product contamination with <u>Klebsiella pneumonia</u>.</i>
Recall Class	Product Identification	Distribution	Affected Dates
N/A	Suspect Lot Number: <ul style="list-style-type: none">• F16091-61 Expiry: 03/2019 <i>See link below for photos</i>	CA , nationwide	Manufactured on: March 3, 2016

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.fda.gov/Safety/Recalls/ucm517443.htm>